Transport packaging

If the device needs to be transported, use the transport packaging supplied. To ensure safe transport, the following precautions must be observed.

1. Screw the adjustable feet as far as possible into the housing.
2. Secure the pendulum.
3. Secure the brace by fastening it to the press station using the threaded bolts on each side.
4. Place the device in the packaging. Make sure that the device is positioned with its feet in the cut-outs prepared for this purpose.
5. Place the cardboard supports evenly along all four sides of the device to create a stable support surface for subsequently placing the lid on top.
6. Make sure that all accessories are fastened.
7. Ship the box on a pallet.

! The pendulum must be locked for any internal transport, especially for shipping. In this case, all additional weights must be removed!
The pendulum must be secured so that it cannot swing freely.

The transport lock on the pendulum must be removed before commissioning.

